

# S.M. GAUGE

## COMPANY

SM GAUGE COMPANY LTD.  
308 / 310 LODGE CAUSEWAY, FISHPONDS, BRISTOL BS16 3RD  
E-mail: sales@pressuregauge.co.uk

Tel: 0117 965 4615  
Fax: 0117 958 3660  
Website: www.pressuregauge.co.uk

### CERTIFICATE OF CALIBRATION

CERTIFICATE NUMBER:- 113584

CUSTOMER:- MR A CHAPMAN

DATE OF TEST:- 01/02/2016

S.M. GAUGE REF:- 277985

SIZE AND TYPE:- 4" PRESSURE GAUGE

STANDARD APPLIED:- BS EN 837-1

RANGE:- 0-460 UNITS:- PSI

TEMPERATURE OF TEST:- 20°C±5°C

SERIAL NUMBER:- 769656

TESTED AGAINST:- 10583260

ACCURACY OF INSTRUMENT:- ± 1% FSD

ADDITIONAL INFO:-

Test Equipment Readings in PSI	Reading of Instrument on Test			
	Rising	Correction	Falling	Correction
100.00	100.00	+0.00	100.00	+0.00
200.00	200.00	+0.00	200.00	+0.00
300.00	296.00	+4.00	296.00	+4.00
400.00	396.00	+4.00	396.00	+4.00
460.00	457.50	+2.50		

Where applicable the instrument was lightly tapped before each reading. Positive corrections are to be added to, and negative corrections subtracted from instrument readings. Figures given under "falling" were taken on fall from maximum scale reading.

The above instrument has been calibrated against equipment traceable to National Standards.

Signed:



Authorised signatory for S.M. Gauge Co.